

Material Safety Data Sheet

Valisafe EN867 Part 5 Helix Indicators
Code/Art No.: V3502034

1. COMPANY IDENTIFICATION

Name of Manufacturer:	Valisafe Ltd
Address:	Valisafe, Twyford Road, Bishops Stortford, CM23 3LJ
Telephone:	+44 (0) 1279 461641
Fax Number:	+44 (0) 1279 461643

2. PRODUCT INFORMATION

Name of Product:	Valisafe EN867 Part 5 Helix Indicator – 3.5 min / 134°C
Code:	V3502034
Size:	Indicator: Length: 73.5mm, Height: 5.0mm
Material:	Strips: Wood free offset cardboard, 170 gram with no laminate; Ink: Lead free.

3. INTENDED USE

Placed in the chamber of a Helix to test if hollow loads are properly sterilized.

Contents:	250 pcs per box
Life time cycle:	3 years
Sterilization method:	Steam
Resterilization:	NA
Latex Free:	Yes
DEHP free:	Yes
CE Class:	I
Standards:	EN ISO 11140-1 (Class 2); EN867 Part 5

4. PRODUCT FAMILY

The Valisafe EN867 Part 5 Helix Indicators are part of the sterilization monitoring family.

5. STORAGE CONDITIONS

Product should be stored in cool, dry place, away from sunlight.	
Temperature range:	10°C - 30°C
Relative humidity:	30 – 70 %



6. TEST RESULTS

When test is performed correctly, indicator should change from pink to brown or darker.

DISCLAIMER

The Information supplied in this document is for guidance only and should not be construed as a warranty. All implied warranties are expressly disclaimed, including without limitation any warranty of merchantability and fitness for use. All users of the material are responsible for ensuring that it is suitable for their needs, environment and end use. All data is subject to change as Valisafe deem appropriate. Refer to www.valisafe.com for more information.